

S/S BOWL WITHOUT SPOUT DIAM 80 MM, 180 ML

Ref. AI99080



25/08/2024 19:13:22 - For information purposes only, can be subject to change

GSH | 30 rue Jean de Guiramand
13290 Aix-en-Provence | France

www.gsh-med.fr

+33(0)4 42 90 31 31

info@gsh-med.fr

S/S BOWL WITHOUT SPOUT DIAM 80 MM, 180 ML



Ref. AI99080

Allows surgical instruments to be stored and transported

Description

Holtex instruments are all made of premium stainless steel (AISI 304, 410, 420 and 420A, depending on the type of instrument) to ensure good durability and corrosion resistance.

The steels come into contact with human tissues, so they must comply with international standards and regulations.

Holtex guarantees you medical devices that comply with Directive 93/42/EC and ISO 13485 (2016).

- In the "ATTACHED DOCUMENTS" tab, you can download the "Cleaning and sterilization" leaflet

Features

CE Marking	CE	Medical device	Class I
Bar code	3661040019864	HS Code	90189090
Warranty (months)	12	Country	Pakistan
DIMENSIONS AND WEIGHT			
Packed unit / Width (cm)	8	Bare Unit / Width (cm)	8
Packed unit / Height (cm)	5	Bare Unit / Height (cm)	5.4
Packed unit / Length (cm)	8	Bare Unit / Length (cm)	8
Packed unit / Weight	0.057	Bare Unit / Weight (kg)	0.057
SPECIFIC STORAGE CONDITIONS			
Minimum temperature	NA	Max temperature	NA
Minimum humidity	NA	Maximum humidity	NA
OTHER FEATURES			
Packed unit / Quantity	1		

25/08/2024 19:13:22 - For information purposes only, can be subject to change

GSH | 30 rue Jean de Guiramand
13290 Aix-en-Provence | France

www.gsh-med.fr

+33(0)4 42 90 31 31

info@gsh-med.fr